



**SDI Review Form 1.6**

Journal Name:	<a href="#">Journal of Pharmaceutical Research International</a>
Manuscript Number:	Ms_JPRI_40460
Title of the Manuscript:	A Novel Stress Indicating RP-HPLC Method Development and Validation for the Simultaneous Estimation of Velpatasvir and Sofosbuvir in bulk and its Tablet Dosage Form
Type of the Article	Original Research Article

**General guideline for Peer Review process:**

This journal's peer review policy states that **NO** manuscript should be rejected only on the basis of '**lack of Novelty**', provided the manuscript is scientifically robust and technically sound. To know the complete guideline for Peer Review process, reviewers are requested to visit this link:

(<http://www.sciencedomain.org/page.php?id=sdi-general-editorial-policy#Peer-Review-Guideline>)

**PART 1: Review Comments**

	Reviewer's comment	Author's comment (if agreed with reviewer, correct the manuscript and highlight that part in the manuscript. It is mandatory that authors should write his/her feedback here)
<b>Compulsory</b> REVISION comments	<p><b>1. Abstract:</b> Line 15: column specification 250 x 2.6 mm, whereas <b>Chromatographic condition &amp; other places:</b> Line 84: column specification 250 x 4.6 mm,</p> <p><b>2. Abstract:</b> Line 18: UV detection 258 nm, whereas <b>Chromatographic condition &amp; other places:</b> Line 91: UV detection 285 nm,</p> <p><b>3. Abstract:</b> Line 23 &amp; 24: LOD &amp; LOQ written as for VEL and SOF were found to be 0.0068-0.104 Åµg/mL and 0.029-0.342 Åµg/mL respectively, whereas <b>Limit of detection (LOD) and Limit of quantification (LOQ):</b> Line 214: LOD &amp; LOQ written as 0.0068µg/ml &amp; 0.029µg/ml for VEL and 0.104 µg/ml &amp; 0.347 µg/ml for SOF respectively,</p> <p><b>4. % Assay:</b> Calculation should be done by the formula =(Weight of substance estimated *100)/weight of substance labelled Line 227: Table 7: results shown are not as per formula.</p>	<p>1. Typhographical mistake and Corrected as 250x 4.6 mm.</p> <p>2. Typhographical mistake, and corrected as 285 nm.</p> <p>3. Corrected LOD and LOQ for VEL 0.0068-0.029 µg/mL and 0.104-0.342 µg/mL for SOF respectively.</p> <p>4. Sorry sir that is some miss calculation with amount found, only i calculate %Assay by taking with peak area.</p>
<b>Minor</b> REVISION comments	1. Reference styles should be uniform for all.	
<b>Optional/General</b> comments	1. Resolution factor calculation may be given with formula and specifications.	